

Tests of medical and personal protection equipment accredited in testing laboratory No. 1004 from 16.6.2020, OA No. 383

Ordinal number	Test procedure / method name	Test procedure / method identification	Tested object
F 8.13	Test method for resistance to ignition	ČSN EN 149+A1, art. 8.6 ČSN EN 136, art. 8.5.1 ČSN EN 13274-4, method 2, 3	PPE – face masks
P 12.5	Testing of medical face masks - Determination of breathability (differential pressure) - construction and design	ČSN EN 14683+AC, Annex C P-20-25 (ČSN EN 14683+AC, art. 5.1)	Medical face masks
P 12.6	Determination of resistance against penetration by synthetic blood	ISO 22609	Medical face masks
P 13.22	Determination of protective properties - Determination of a protected area of a face - Determination of protection against drops and splashes of liquids	ČSN EN 168 Art. 10.2 Art. 12	PPE – personal eye-protection
P 13.23	Determination of resistance to penetration by a spray of liquid (spray test) after a practical test by wearing	ČSN EN ISO 17491-4 ČSN EN 13034+A1, Art. 5.2 ČSN EN 14605+A1, Art. 4.3.4	PPE – protective clothing
P 13.24	Determination of resistance to penetration by a jet of liquid (jet test) after a practical test by wearing	ČSN EN ISO 17491-3 ČSN EN 14605+A1, Art. 4.3.4	PPE – protective clothing