



NOTIFIED BODY No. 1023
Institute for Testing and Certification, Inc., Zlín, Czech Republic

EC CERTIFICATE

No. 10 0301 QS/NB

issued in compliance with the Council Directive 93/42/EEC as amended, which is implemented by the Czech Government Order No. 336/2004 (Collection of Laws), certifies that the medical devices of Class IIb:

Anesthesia machine

Models: PAS-200A, PAS-200B, PAS-200C, PAS-200D, PAS-200E

Ventilator

Models: WHYP-100, WHYH-100A, WHYH-100B, WHYH-100C

manufactured by company

Poweam Medical Co., Ltd

No.580 Moling Street Nanjing City, Jiangsu, China


is manufactured under conditions fulfilling the quality system requirements of Annex II, Section 3.2 of the Directive 93/42/EEC, as amended.

The Notified Body No. 1023 has performed an audit of the above products quality system covering the design, manufacture and final inspection of the certified products. The quality system has been assessed, approved and is subject to continuous surveillance according to Annex II, Sections 3.3 and 5 of the Directive 93/42/EEC. The detailed description of the system parts, requirements and measures applied by the manufacturer are presented in the Final Report No. 803601115/2011, which is enclosed to this Certificate.

This Certificate is issued under the following conditions:

1. It applies only to the quality system maintained in the manufacture of above referenced models of medical devices and it does not substitute the design or type-examination procedures, if requested.
2. The Certificate remains valid until the manufacturing conditions or the quality system are changed but until the 14th April 2020 at the latest.
3. The Certificate validity is conditioned by positive results of surveillance audits.
4. After fulfilling the relevant EU legislation, the manufacturer shall affix to each medical device, of the above referenced models, the CE-marking followed by the number of the Notified Body according to this example:

CE 1023



Issued in Zlín, on 15th April 2015

RNDr. Radomír Čevelík

Representative of the Notified Body No. 1023