



Notified Body No 1023  
**INSTITUTE FOR TESTING AND CERTIFICATION, Inc.**  
Zlín, Czech Republic – [www.itczlin.cz](http://www.itczlin.cz)

# EC CERTIFICATE

## No. 10 0649 QS/NB/b

issued in compliance with the Council Directive 93/42/EEC as amended, which is implemented by the Czech Government Order No. 336/2004 (Collection of Laws), certifies that the medical devices of Class IIb:

### 1) LFS Diode Laser Treatment System

**models:** DM-800, DM-810, DM-830, DM-880A, DM-820, DM-860  
DM-808A, DM-808B, DM-808C, DM-808.

### 2) Fractional Laser Treatment System

**models:** DM-830, DM-840, DM-850, DM-870, DM-880

manufactured by company

### **Beijing LefisTechnology Co.,Ltd.**

**Room 054,2/F,Building1,No.3 Liangshuihe North ,Xiaohongmenxiang Santaishan Road,Chaoyang District,Beijing,China(Mainland)**

are manufactured under conditions fulfilling the quality system requirements of Annex II, Section 3.2 of the Directive 93/42/EEC.


The Notified Body No. 1023 has performed an audit of the above products quality system. The quality system has been assessed, approved and is subject to continuous surveillance according to Annex II, Sections 3.3 and 5 of the Directive 93/42/EEC. The detailed description of the system parts, requirements and measures applied by the manufacturer are presented in the Final Report No. 803600922/2010 and 343601122, which are enclosed to this Certificate.

#### *Condition of this Certificate use and related information:*

1. *It applies only to the quality system maintained in the manufacture of above referenced models of medical devices and it does not substitute the design or type-examination procedures, if requested.*
2. *The Certificate remains valid until the manufacturing conditions or the quality system are changed but until the 28<sup>th</sup> July, 2017 at the latest.*
3. *The Certificate validity is conditioned by positive results of surveillance audits.*
4. *After fulfilling the relevant EU legislation, the manufacturer shall affix to each medical device, of the above referenced models, the CE-marking followed by the number of the Notified Body according to this example:*

**CE 1023**



  
RNDr. Radomír Čevelík  
Representative of the Notified Body No. 1023

Issued in Zlín, on 8<sup>th</sup> December 2013  
Replaces withdrawn Certificate No. 10 0649 QS/NB/a issued on 18<sup>th</sup> November 2012